



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Louis H. Plunkett III is a certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedures are a full, true, and correct copy of the original.



Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 25th day of September, 2018

by Louis Plunkett

Shawna Mcknight
NOTARY PUBLIC

9/25/2018
DATE

Commission expires:



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
 TDD 855-4114
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov



**ARIZONA DEPARTMENT OF PUBLIC SAFETY
CRIMINAL JUSTICE SUPPORT DIVISION
INTOXILYZER SERVICE REQUEST**

Admin Use Only	
CR#	
Date:	
Initials:	

<input type="checkbox"/> Repair	Date Submitted	Intoxilyzer Serial Number
<input checked="" type="checkbox"/> Preventative Maintenance	Sep 25, 2018	80-0033356

Submitting Agency Name	Location of Equipment/Agency
LAKE HAVASU CITY PD	2360 MCCULLOCH BLVD N, LAKE HAVASU CITY, AZ 86403

Last Name	First Name	Badge
PLUNKETT III	LOUIS	253

Cell or Primary Phone	Work Phone	Email
	928-680-5422	plunkettl@hcaz.gov

Other Equipment Sent:

Gas Cylinder

Power Cord

Problem Description (if applicable) (please be specific)

THIS UNIT IS SCHEDULED FOR ITS ANNUAL PM.

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS H. BLUNKETT III AGENCY MACPD

DATE 09-26-18 TIME 0707

INTOXILYZER SERIAL # 80-003366 LOCATION MACPD

(X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

(X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(X) 4. Air blank completed.

(X) 5. Calibration check completed. Test results 0. 098 AC.

(X) 6. Air blank completed.

(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8100
Location: MAIN JAIL
Serial Number: 90-003356
Cone Version: 8105.46
09/05/2013 07:07:42

Standard Lot#: 139171094
Last Changed By: J. SOMORDEL #187

OPS: L. PLUNKETT 111 #053
LAKE HAVASU CITY PD

Operator: L. PLUNKETT 111 #053
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	07:08:13
0.100 Cal Check	0.098	07:08:34
Air Blank	0.000	07:09:03

Preventative Maintenance
Performed -- Timer Reset

143
263

Exhibit G-4. Standard Quality Assurance Procedure Intoxilyzer, Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOUIS W. BLUNKETT III AGENCY LACPD

DATE 09-25-18 TIME 0711

INTOXILYZER SERIAL # 80-003356 LOCATION LACPD

(X) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (X) 1. Clock time check.
(X) 2. Date check.

OPERATIONAL TESTS

- (X) 1. Alcohol-free subject test result 0.000 AC.
(X) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(X) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(X) 4. Standard Calibration Check standard 0.100 AC. Result 0.098 AC.

Instrument is operating properly and accurately. YES X NO

COMMENTS G-2/G-4/G-2 DONE FOR ANNUAL PM
GOT REF DURING DEFICIENT SAMPLE TEST
IMPROPER SAMPLE DURING AIR BLANK

SIGNATURE [Signature] 253

DPS Form Exh G-4 (Rev 05-01)

Historical Note

New Exhibit G-4 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INSTRUMENT #000
 Location: MAIN LAB
 Serial Number: 80-000056
 Core version: 3105.48
 09/25/2018 07:11:00

Standard Lot#: 1397100A
 Last Changed By: J. SONCROU #187

OAS: L. PLUNKETT 111 #250
 LAKE HAUBAU CITY PD

Operator: L. PLUNKETT 111 #250
 LAKE HAUBAU CITY PD

Subject: SOAP,TEST
 DOB: 01/01/2000
 Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/dL	Time
Air Blank	0.000	07:12:10
Diagnostic Test	Pass	07:12:48
Air Blank	0.000	07:13:15
0.100 Cal Check	0.098	07:13:36
Air Blank	0.000	07:14:06
Subject Test	0.000	07:14:33
Air Blank	ABT*	07:14:40
Air Blank	IFS**	07:14:47

*Sequence Aborted
 **Improper Sample

Not a Successfully
 Completed
 Test Sequence **493**
253

INSTRUMENT #000
 Location: MAIN LAB
 Serial Number: 80-000056
 Core version: 3105.48
 09/25/2018 07:26:00

Standard Lot#: 1397100A
 Last Changed By: J. SONCROU #187

OAS: L. PLUNKETT 111 #250
 LAKE HAUBAU CITY PD

Operator: L. PLUNKETT 111 #250
 LAKE HAUBAU CITY PD

Subject: SOAP,TEST
 DOB: 01/01/2000
 Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/dL	Time
Air Blank	0.000	07:29:10
Diagnostic Test	Pass	07:29:47
Air Blank	0.000	07:29:14
0.100 Cal Check	0.098	07:29:35
Air Blank	0.000	07:30:15
Subject Test	0.00*	07:30:12
Air Blank	0.000	07:30:45
Air Blank	0.000	07:31:12
Subject Test	0.00**	07:31:37
Air Blank	RF***	07:31:43

*Deficient Sample
 **Invalid Sample
 Begin new deprivation period
 ***RF - Detect

Not a Successfully
 Completed
 Test Sequence

493
253

INTOXILYZER 8000
Location: MAIN LAB
Serial Number: 80-003356
Core Version: 8105.08
09/25/2018 07:17:25

Standard Lot#: 139.7100A1
Last Changed By: J. BONDRU #187

Ops: L. PLUNKETT (11) #253
LAKE HARVEST CITY PD

Operator: L. PLUNKETT (11) #253
LAKE HARVEST CITY PD

Subject: SQAP, TEST
DOB: 11/01/2000
Sex: M Weight: 200

15 Minute Dehydration Period: Yes

Test	g/210L	Time
Air Blank	0.000	07:18:17
Diagnostic Test	Pass	07:19:04
Air Blank	0.000	07:19:22
1.100 Cal Check	0.098	07:19:40
Air Blank	0.000	07:20:12
Subject Test	RF1*	07:21:01
Air Blank	0.000	07:21:54

*RF1 Detect

Not a Successfully
Completed
Test Sequence

LP3
263

INTOXILYZER 8000
Location: MAIN LAB
Serial Number: 80-003356
Core Version: 8105.08
09/25/2018 07:22:57

Standard Lot#: 139.7100A1
Last Changed By: J. BONDRU #187

Ops: L. PLUNKETT (11) #253
LAKE HARVEST CITY PD

Operator: L. PLUNKETT (11) #253
LAKE HARVEST CITY PD

Subject: SQAP, TEST
DOB: 11/01/2000
Sex: M Weight: 200

15 Minute Dehydration Period: Yes

Test	g/210L	Time
Air Blank	0.000	07:23:27
Diagnostic Test	Pass	07:24:04
Air Blank	0.000	07:24:31
1.100 Cal Check	0.098	07:24:52
Air Blank	RF5*	07:25:21
Air Blank	0.000	07:25:50

*RF5 Detect

Not a Successfully
Completed
Test Sequence

LP3
263

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS H. PLUNKETT, III AGENCY LHCPD

DATE 09-25-18 TIME 0735

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD

(X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 097 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  JF 255

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000
Location: MAIN DEPT
Serial Number: 60-113356
Date Version: 8105148
09/25/2018 07:35:38

Standard Lot#: 39710A
Last Changed By: J. SONDROL #187

Operator: L. PUNKETT (1) #253
LAKE HAVASU CITY PD

Operator: L. PUNKETT (1) #253
LAKE HAVASU CITY PD

Test	g/2 DL	Time
Air Blank	0.000	07:36:08
0.100 Cell Check	0.097	07:36:29
Air Blank	0.000	07:36:58

Preventative Maintenance
Performed -- Timer Reset

LP3
263