



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Michael T. Maloy, Jr. is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Michael T Maloy Jr #1281
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)


Subscribed and sworn to before me this 28th day of September.

2022 by Michael T Maloy Jr

[Signature]
NOTARY PUBLIC

2-26-23
DATE

Commission expires:



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD
DATE 09-28-2022 TIME 1632
INTOXILYZER SERIAL # 90-003680

1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- 1. Clock time check
- 2. Date check

OPERATIONAL TESTS

- 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- 5. Standard Calibration Check:
0.095 AC
- 6. Air Blanks Completed
- 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES NO

COMMENTS:

SIGNATURE Michael Maloy Jr #281

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
 Serial Number: 90-003680
 Software Version: 9439.01.00
 Analytical Report Number: 36800928220023

QAS: MALOY, MICHAEL
 QAS Permit #: 30178
 Agency: LAKE HAVASU PD
 Last 31-Day Check: 08/28/2022
 Last Annual Maintenance: 11/15/2021

SUBJECT INFORMATION

Name: 31-Day Check
 Test Date: 09/28/2022
 Driver's License #:
 Gender:
 Date of Birth:
 Age:

Weight:
 State of Issue:
 Driver's License Expiration:
 Deprivation Start Time: 16:00
 15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MALOY, MICHAEL
 Agency: LAKE HAVASU PD

Permit #: 30178

STANDARD INFORMATION

Standard Value: 0.100
 Standard Lot #: 23621100A4
 Expiration Date: 11/05/2023
 Bottle #: 003
 Last Changed By: MALOY, MICHAEL

Test	g/210L	Time
Air Blank	0.000	16:32:40
Diagnostic Test	PASS	16:33:16
Air Blank	0.000	16:33:53
Calibration Chk	0.095	16:34:15
Air Blank	0.000	16:34:55
Subject Test 1	DEF*	16:38:08
Air Blank	0.000	16:38:53
Air Blank	0.000	16:39:34
Subject Test 2	0.000	16:40:05
Air Blank	0.000	16:41:19
Wait		16:45:22
Air Blank	0.000	16:45:59
Subject Test 3	INV**	16:46:44
Air Blank	RFI***	16:46:57

Permit #: 30178

RESULTSEXCEPTION MESSAGES

*Deficient Sample
 **Invalid Sample - Begin new deprivation period
 ***RFI Detect

* Deficient Sample
 ** Invalid Sample - Begin new deprivation period
 *** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
 Consecutive subject test results must not differ by more than 0.020 g/210L.
 Standard check results must be $\pm 10\%$.*