



EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LHCPD  
DATE 04-14-2020 TIME 2214  
INTOXILYZER SERIAL # 80-003356 LOCATION SIGN 8

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 099 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy Jr #281

INTOXILYZER 8000  
Location: CHANNEL SIGN 8  
Serial Number: 80-003356  
Core Version: 8105.48  
04/14/2020 22:14:58

-----  
Standard Lot#: 35019100A1  
Last Changed By: M. MALOY JR #281  
-----

QAS: M. MALOY JR #281  
LAKE HAVASU CITY PD  
-----

Operator: M. MALOY JR #281  
LAKE HAVASU CITY PD  
-----

Test	g/210L	Time
Air Blank	0.000	22:15:54
0.100 Cal Check	0.099	22:16:14
Air Blank	0.000	22:16:44

Preventative Maintenance  
Performed -- Timer Reset

281

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LULCPD  
DATE 04-14-2020 TIME 2224  
INTOXILYZER SERIAL # 80-003356 LOCATION SIGN 8

(  ) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  ) 1. Clock time check.
- (  ) 2. Date check.

OPERATIONAL TESTS

- (  ) 1. Alcohol-free subject test result 0. 000 AC.
- (  ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- (  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- (  ) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

CHANGED GAS STANDARDS  
NEW LOT # 35019100A1  
EXP DATE: 02-05-2022

SIGNATURE Michael Maloy Jr #281



INTOXILYZER 8000

Location: CHANNEL SIGN 8

Serial Number: 80-003356

Core Version: 8105.48

04/14/2020

22:24:27

Standard Lot#: 35019100A1

Last Changed By: M. MALOY JR #281

QAS: M. MALOY JR #281

LAKE HAVASU CITY PD

Operator: M. MALOY JR #281

LAKE HAVASU CITY PD

Subject: SQAP, TEST

DOB: 01/01/2000

Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:26:03
Diagnostic Test	Pass	22:26:38
Air Blank	0.000	22:27:05
0.100 Cal Check	0.099	22:27:26
Air Blank	0.000	22:27:56
Subject Test	0.000	22:28:22
Air Blank	ABT*	22:28:28
Air Blank	IPS**	22:28:37

\*Sequence Aborted

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
 Location: CHANNEL SIGN 8  
 Serial Number: 80-003356  
 Core Version: 8105.48  
 04/14/2020 22:41:49

---

Standard Lot#: 35019100A1  
 Last Changed By: M. MALOY JR #281

---

QAS: M. MALOY JR #281  
 LAKE HAVASU CITY PD

---

Operator: M. MALOY JR #281  
 LAKE HAVASU CITY PD

---

Subject: SQAP,TEST  
 DOB: 01/01/2000  
 Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	22:42:42
Diagnostic Test	Pass	22:43:17
Air Blank	0.000	22:43:44
0.100 Cal Check	0.098	22:44:05
Air Blank	0.000	22:44:35
Subject Test	0.000*	22:47:42
Air Blank	0.000	22:48:10
Air Blank	0.000	22:48:37
Subject Test	XXX**	22:48:52
Air Blank	RFI***	22:49:04

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

\*\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. MALOY #281 AGENCY LUICPD  
DATE 04-14-2020 TIME 2252  
INTOXILYZER SERIAL # 80-003356 LOCATION S16N 8

- (  ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (  ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (  ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (  ) 4. Air blank completed.
- (  ) 5. Calibration check completed. Test results 0. 096 AC.
- (  ) 6. Air blank completed.
- (  ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Michael Maloy #281

281

INTOXILYZER 8000  
Location: CHANNEL SIGN 8  
Serial Number: 80-003356  
Core Version: 8105.48  
04/14/2020 22:52:06

-----  
Standard Lot#: 35019100A1  
Last Changed By: M. MALOY JR #281  
-----

QAS: M. MALOY JR #281  
LAKE HAVASU CITY PD  
-----

Operator: M. MALOY JR #281  
LAKE HAVASU CITY PD  
-----

Test	g/210L	Time
Air Blank	0.000	22:52:36
0.100 Cal Check	0.096	22:52:57
Air Blank	0.000	22:53:26

Preventative Maintenance  
Performed -- Timer Reset





7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

## Certificate of Analysis

Certificate ID: 12621  
App #: BAC105L100T  
Cylinder Size: 105L  
Lot Number: 35019100A1  
Expiration: 2/5/2022

1000 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Content: 10.35 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Alc	1000 ppm	+/-0.002 BAC(G/210L)	NDIR
Moisture	Balance	[5.2 ppm]	

**Distributed by:**

CMI Inc.  
316 East Ninth Street  
Owensboro, KY 42303  
Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)

Available in:  
Certified Reference Material - 262.4 µmol/mol  
Cylinder Filling - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Analyst: [Signature]  
Date: [Signature]

01-06-2020  
Issuance Date



All calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items mentioned in this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular application. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2005 Accredited Laboratory