



POLICE DEPARTMENT
LAKE HAVASU CITY
2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Josh Sondrol
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)


Subscribed and sworn to before me this 1 day of June

2016 by Josh Sondrol

Carla Duarte
NOTARY PUBLIC

3-29-17
DATE

Commission expires:



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
TDD 855-4114
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

7101

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD

DATE 6/1/16 TIME 0211

INTOXILYZER SERIAL # 80-003355 LOCATION EB BOOKING

() 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

187

INTOXILYZER 8000

Location: EAST BANK BOOKING

Serial Number: 80-003355

Core Version: 8105.48

06/01/2016

02:11:34

Standard Lot#: 20314100A3

Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187

LAKE HAVASU CITY PD

Operator: J. SONDROL #187

LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	02:12:15
0.100 Cal Check	0.100	02:12:36
Air Blank	0.000	02:13:05

Preventative Maintenance
Performed -- Timer Reset

2181

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. Sondreal 187 AGENCY LNC PD
DATE 6/1/16 TIME 0214
INTOXILYZER SERIAL # 80-003355 LOCATION EB BOOKING

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE J. Sondreal 187

Y0181

INTOXILYZER 8000
Location: EAST BANK BOOKING
Serial Number: 80-003355
Core Version: 8105.48
06/01/2016 02:14:00

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	02:14:47
Diagnostic Test	Pass	02:15:22
Air Blank	0.000	02:15:49
0.100 Cal Check	0.100	02:16:10
Air Blank	0.000	02:16:39
Subject Test	0.000	02:17:25
Air Blank	ABT*	02:17:31
Air Blank	IPS**	02:17:39

*Sequence Aborted
**Improper Sample

Not a Successfully Completed Test Sequence

7101

INTOXILYZER 8000
Location: EAST BANK BOOKING
Serial Number: 80-003355
Core Version: 8105.48
06/01/2016 02:18:17

Standard Lot#: 20314100A3
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	02:19:06
Diagnostic Test	Pass	02:19:41
Air Blank	0.000	02:20:08
0.100 Cal Check	0.100	02:20:29
Air Blank	0.000	02:20:59
Subject Test	0.181*	02:24:06
Air Blank	0.000	02:24:39
Air Blank	0.000	02:25:06
Subject Test	XXX**	02:25:41
Air Blank	RFI***	02:25:48

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

7/187

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDREL 187 AGENCY LNC PD

DATE 6/1/16 TIME 0226

INTOXILYZER SERIAL # 80-003355 LOCATION EB BOOKING

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record/ Attach the record to the completed checklist.

SIGNATURE J Sondrel 187

J187

INTOXILYZER 8000
Location: EAST BANK BOOKING
Serial Number: 80-003355
Core Version: 8105.48
06/01/2016 02:26:38

Standard Lot#: 20314100A3
Last Changed By: J. SONDRÖL #187

QAS: J. SONDRÖL #187
LAKE HAVASU CITY PD

Operator: J. SONDRÖL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	02:27:06
0.100 Cal Check	0.099	02:27:27
Air Blank	0.000	02:27:57

Preventative Maintenance
Performed -- Timer Reset