



EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDRAL 187 AGENCY LHC PD  
DATE 3/23/16 TIME 1723  
INTOXILYZER SERIAL # 80-001029 LOCATION LB CHANNEL SIGN 8

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondral 187

DPS Form Exh G-2 (Rev 05-01)

*JSR*

INTOXILYZER 8000  
Instrument Initialization  
05:55:40 03/23/2016

INTOXILYZER 8000  
Location: LB CHANNEL SIGN 8  
Serial Number: 80-001029  
Core version: 8:15.46  
03/23/2016 17:23:06

-----

Standard Lot#: 223:5100A  
Last Changed By: J. SONDRAL #187

-----

DAS: J. SONDRAL #187  
LAKE HAVASU CITY PD

-----

Operator: J. SONDRAL #187  
LAKE HAVASU CITY PD

-----

Test	g/210L	Time
Air Blank	0.000	17:23:41
0.100 Cal Check	0.100	17:24:02
Air Blank	0.000	17:24:31

Preventative Maintenance  
Performed -- Timer Reset

**EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**

**ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000**

**STANDARD QUALITY ASSURANCE PROCEDURE**

QA SPECIALIST J SONDROL 187 AGENCY LHC PD  
 DATE 3/23/16 TIME 1725  
 INTOXILYZER SERIAL # 80-001029 LOCATION LB CHARGER SIGN 8

() 1. Display Reads "PUSH BUTTON TO START"

**DIAGNOSTIC TESTS**

- () 1. Clock time check.
- () 2. Date check.

**OPERATIONAL TESTS**

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.  
 Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.  
 Not a Successfully Completed Test Sequence printed  
 Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES  NO

**COMMENTS**

SIGNATURE J SONDROL 187

INTOXILYZER # 8000  
 Location: LB CHARGER SIGN 8  
 Serial Number: 80-001029  
 Core Version: 8.05.49  
 03/23/2016 J187 17:25:31

Standard Lot#: 22315100A1  
 Last Changed By: J. SONDROL #187

DAS: J. SONDROL #187  
 LAKE HAVASU CITY PD

Operator: J. SONDROL #187  
 LAKE HAVASU CITY PD

Subject: SGAP, TEST  
 DOB: 01/01/1950  
 Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:26:18
Diagnostic Test	Pass	17:26:55
Air Blank	0.000	17:27:23
0.100 Cal Check	1.100	17:27:43
Air Blank	0.000	17:28:13
Subject Test	0.000	17:28:32
Air Blank	RST*	17:28:35
Air Blank	IPS**	17:28:46

\*Sequence Aborted  
 \*\*Improper Sample

**Not a Successfully  
Completed  
Test Sequence**

DPS Form Exh G-4 (Rev 05-01)

*(Faint mirrored text from reverse side of page)*

Test	g/210L	Time
Air Blank	0.000	17:26:20
Diagnostic Test	Pass	17:26:58
Air Blank	0.000	17:27:25
0.100 Cal Check	1.100	17:27:46
Air Blank	0.000	17:28:15
Subject Test	0.000	17:28:23
Air Blank	0.000	17:28:56
Air Blank	0.000	17:29:25
Subject Test	0.000	17:29:53
Air Blank	0.000	17:30:10

\*Sequence Aborted  
 \*\*Improper Sample  
 Best 15 min deprivation period  
 (RST) Detect

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SANDROGL 187 AGENCY LHC PD  
DATE 3/23/16 TIME 1740  
INTOXILYZER SERIAL # 80-001029 LOCATION LB CHANNEL SIGN 8

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sandrogl 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000  
Location: LB CHANNEL SIGN 8  
Serial Number: 80-001029  
Core Version: 8105.48  
03/23/2016 17:40:12  
Standard Lot#: 22315100A1  
Last Changed By: J. SANDROGL #187  
DAS: J. SANDROGL #187  
LAKE HAVASU CITY PD  
Operator: J. SANDROGL #187  
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	17:40:42
0.100 Cal Check	0.099	17:41:02
Air Blank	0.000	17:41:32

Preventative Maintenance  
Performed -- Timer Reset