




POLICE DEPARTMENT
LAKE HAVASU CITY

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.




Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 4 day of May .


 2016 by JOSH SONDROL



NOTARY PUBLIC

 5-4-16
DATE

Commission expires:



Area Code 928

Administration 855-4884
Business Office 855-1171
Investigations 855-5775
Patrol 855-0515

Fax #'s
Administration 680-5430
Business Office 680-5431
Investigations/Patrol 680-5432

Emergency 9-1-1
Non-Emergency Dispatch/ 855-4111
 TDD 855-4114
Havasu Silent Witness 854-TIPS

E-mail: police@lhcaz.gov

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD
DATE 5/3/16 TIME 2309
INTOXILYZER SERIAL # 80-003355 LOCATION LB CHANNEL SIGN 8

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Instrument Initialization
22:44:06 05/03/2016

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 9105.48
05/03/2016 23:09:43

Standard Lot#: 2031410093
Last Changed By: J. SONDROL #187

OAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	23:10:25
0.100 Cal Check	0.100	23:10:45
Air Blank	0.000	23:11:15

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD
 DATE 5/3/16 TIME 2312
 INTOXILYZER SERIAL # 80-003355 LOCATION LB CHANNEL SIGN 8

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
 () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0.000 AC.
 () 2. Error recognition logic system functioning.
 Not a Successfully Completed Test Sequence printed
 () 3. Proper sample recognition system.
 Not a Successfully Completed Test Sequence printed
 Deficient sample printed.
 () 4. Standard Calibration Check standard 0.100 AC. Result 0.100 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE J Sondrol

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000
 LOCATION: LB CHANNEL SIGN 8
 SERIAL NUMBER: 80-003355
 CORE VERSION: 8105.48
 05/03/2016 23:17:10

STANDARD LOT#: 20314100-F3
 LAST CHANGED BY: J. SONDROL #187

QAS: J. SONDROL #187
 LAKE HAVASU CITY PD

OPERATOR: J. SONDROL #187
 LAKE HAVASU CITY PD

SUBJECT: SOAP TEST
 DOB: 01/01/1950
 SEX: M HEIGHT: 123

15 Minute Deprivation Period: Yes

INTOXILYZER 8000
 Location: LB CHANNEL SIGN 8
 Serial Number: 80-003355
 Core Version: 8105.48
 05/03/2016 23:12:59

Standard Lot#: 20314100-F3
 Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
 LAKE HAVASU CITY PD

Operator: J. SONDROL #187
 LAKE HAVASU CITY PD

Subject: SOAP TEST
 DOB: 01/01/1950
 Sex: M Height: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:13:47
Diagnostic Test	Pass	23:14:24
Air Blank	0.000	23:14:52
0.100 Cal Check	0.100	23:15:13
Air Blank	0.000	23:15:42
Subject Test	0.000	23:16:03
Air Blank	ABT*	23:16:09
Air Blank	IPS**	23:16:17

*Sequence Aborted
 **Improper Sample

**Not a Successfully
 Completed
 Test Sequence**

Test	g/210L	Time
Air Blank	0.000	23:17:46
Diagnostic Test	Pass	23:18:23
Air Blank	0.000	23:18:51
0.100 Cal Check	0.100	23:19:11
Air Blank	0.000	23:19:41
Subject Test	0.334*	23:20:48
Air Blank	0.000	23:21:21
Air Blank	0.000	23:23:49
Subject Test	K00**	23:24:15
Air Blank	0.000	23:24:23

x Deficient Sample
 ** Improper Sample
 Begin 15 min. deprivation period
 ***** Detect

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD

DATE 5/3/16 TIME 2325

INTOXILYZER SERIAL # 80-003355 LOCATION LB CHANNEL SIGN 8

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.099 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: LB CHANNEL SIGN 8
Serial Number: 80-003355
Core Version: 8105.48
05/03/2016 23:25:48

Standard Lot#: 29314:00A3
Last Changed By: J. SONDROL #187

GAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.003	23:26:17
0.100 Cal Check	0.099	23:26:38
Air Blank	0.000	23:27:17

Preventative Maintenance
Performed -- Timer Reset