



POLICE DEPARTMENT

2360 McCULLOCH BOULEVARD NORTH  
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Charles Trescher is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

*Charles Trescher*

Signature

STATE OF ARIZONA    )  
  ) ss  
COUNTY OF MOHAVE  )

Subscribed and sworn to before me this First day of October.

. 2016 by

*Jessica Sossaman*  
\_\_\_\_\_  
NOTARY PUBLIC

10/11/2016  
\_\_\_\_\_  
DATE

Commission expires:



Area Code 928

Administration .....	855-4884	Fax #'s	Emergency .....	9-1-1
Business Office .....	855-1171	Administration .....	Non-Emergency Dispatch/ .....	855-4111
Investigations .....	855-5775	Business Office .....	TDD	855-4114
Patrol .....	855-0515	Investigations/Patrol .....	Havasu Silent Witness .....	854-TIPS

E-mail: [police@lhcaz.gov](mailto:police@lhcaz.gov)



C1  
\*219

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST V. TERRAZA AGENCY LAKE MARSH CITY POLICE

DATE 10-2-14 TIME 1747 HRS

INTOXILYZER SERIAL # 18 00078 LOCATION LAC. PD JAIL

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS


- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 000 AC. Result 0. 000 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE



219

INTOXILYZER 8000  
Location: LHC MAIN JAIL  
Serial Number: 80 001873  
Core Version: 8105.48  
10/01/2016

17:43:30

Standard Lot#: 22315100A  
Last Changed By: J. SONDRON #187

QAS: C. TRESCHER #219  
LAKE HAVASU CITY PD

Operator: C. TRESCHER #219  
LAKE HAVASU CITY PD

Subject: SOAP, TEST  
DOB: 10/19/1978  
Sex: M Weight: 220

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:44:40
Diagnostic Test	Pass	17:45:15
Air Blank	0.000	17:45:42
0.100 Cal Check	0.098	17:46:03
Air Blank	0.000	17:46:33
Subject Test	0.000	17:46:52
Air Blank	ABT*	17:46:59
Air Blank	IPS**	17:47:07

\*Sequence Aborted

\*\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: LHC MAIN JAIL  
Serial Number: 80-001873  
Core Version: 8105.48  
10/01/2016

17:43:30

---

Standard Lot#: 22315100A1  
Last Changed By: J. SONDROL #187

---

QAS: C. TRESCHER #219  
LAKE HAVASU CITY PD

---

Operator: C. TRESCHER #219  
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST  
DOB: 10/19/1978  
Sex: M Weight: 220

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:44:40
Diagnostic Test	Pass	17:45:15
Air Blank	0.000	17:45:42
0.100 Cal Check	0.098	17:46:03
Air Blank	0.000	17:46:33
Subject Test	0.000	17:46:52
Air Blank	ABT*	17:46:59
Air Blank	IPS**	17:47:07

\*Sequence Aborted

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: LHC MAIN JAIL  
Serial Number: 80 001873  
Core Version: 8105.48  
10/01/2016

17:43:30

---

Standard Lot#: 22315100A1  
Last Changed By: J. SONDROL #187

---

QAS: C. TRESCHER #219  
LAKE HAVASU CITY PD

---

Operator: C. TRESCHER #219  
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST  
DOB: 10/19/1978  
Sex: M Weight: 220

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:44:40
Diagnostic Test	Pass	17:45:15
Air Blank	0.000	17:45:42
0.100 Cal Check	0.098	17:46:03
Air Blank	0.000	17:46:33
Subject Test	0.000	17:46:52
Air Blank	ABT*	17:46:59
Air Blank	IPS**	17:47:07

\*Sequence Aborted  
\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: LHC MAIN JAIL

Serial Number: 80-001873

Core Version: 8105.48

10/01/2016

17:47:49

---

Standard Lot#: 22315100A1

Last Changed By: J. SONDROL #187

---

QAS: C. TRESCHER #219

LAKE HAVASU CITY PD

---

Operator: C. TRESCHER #219

LAKE HAVASU CITY PD

---

Subject: SQAP,TEST

DOB: 10/19/1978

Sex: M Weight: 220

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:48:46
Diagnostic Test	Pass	17:49:21
Air Blank	0.000	17:49:49
0.100 Cal Check	0.098	17:50:09
Air Blank	0.000	17:50:39
Subject Test	0.031*	17:53:46
Air Blank	0.000	17:54:17
Air Blank	0.000	17:54:45
Subject Test	XXX**	17:55:13
Air Blank	RFI***	17:55:27

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

\*\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: LHC MAIN JAIL  
Serial Number: 80-001873  
Core Version: 8105.48  
10/01/2016

17:47:49

---

Standard Lot#: 22315100A1  
Last Changed By: J. SONDROL #187

---

QAS: C. TRESCHER #219  
LAKE HAVASU CITY PD

---

Operator: C. TRESCHER #219  
LAKE HAVASU CITY PD

---

Subject: SQAP, TEST  
DOB: 10/19/1978  
Sex: M Weight: 220

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:48:46
Diagnostic Test	Pass	17:49:21
Air Blank	0.000	17:49:49
0.100 Cal Check	0.098	17:50:09
Air Blank	0.000	17:50:39
Subject Test	0.031*	17:53:46
Air Blank	0.000	17:54:17
Air Blank	0.000	17:54:45
Subject Test	XXX**	17:55:13
Air Blank	RFI***	17:55:27

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

\*\*\*RFI Detect

**Not a Successfully Completed Test Sequence**



INTOXILYZER 8000  
Location: LHC MAIN JAIL  
Serial Number: 80-001873  
Core Version: 8105.48  
10/01/2016

17:47:49

---

Standard Lot#: 22315100A1  
Last Changed By: J. SONDRON #187

---

QAS: C. TRESCHER #219  
LAKE HAVASU CITY PD

---

Operator: C. TRESCHER #219  
LAKE HAVASU CITY PD

---

Subject: SQAP, TEST  
DOB: 10/19/1978  
Sex: M Weight: 220

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:48:46
Diagnostic Test	Pass	17:49:21
Air Blank	0.000	17:49:49
0.100 Cal Check	0.098	17:50:09
Air Blank	0.000	17:50:39
Subject Test	0.031*	17:53:46
Air Blank	0.000	17:54:17
Air Blank	0.000	17:54:45
Subject Test	XXX**	17:55:13
Air Blank	RFI***	17:55:27

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

\*\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. T. [unclear] AGENCY [unclear]

DATE [unclear] TIME [unclear]

INTOXILYZER SERIAL # [unclear] LOCATION [unclear]

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution [unclear] AC.  
OR  
Pour a standard alcohol concentration solution [unclear] AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. [unclear] AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

19

LINE NUMBER: 837  
DESCRIPTION: 1000 MAIN TATE  
SERIAL NUMBER: 1000000000  
GRAB VALUE: 400.00

17:07:13

DEMAND: 1000000000  
LAST DEMAND BY: 1000000000 #187

MAINT. PERSONNEL #319  
LAST MAINT. CITY: PD

OPERATOR: 1000000000 #210  
LAST MAINT. CITY: PD

LINE	DESCRIPTION	Time
837	Air Blank	17:07:33
838	1000000000	17:07:54
839	Air Blank	17:08:24

Preventative Maintenance  
performed. Dimer Reset