

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SANDRICK 187 AGENCY LNC PD
DATE 9/2/16 TIME 0024
INTOXILYZER SERIAL # 80-001873 LOCATION MARIN JAIL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 097 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE C Sandrick 187

48187

INTOXILYZER 8000
Location: 3811 MAIN TA1L
Serial Number: 80 001873
Firm Version: 8105.4A
09/02/2016 00:24:40

Standard Lot #: 22310790A
Lab Changed By: J. BONDROL #187

QA#: J. BONDROL #187
LAKE HAVASU CITY PD

Operator: J. BONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	00:25:11
0.100 Cal Check	0.097	00:25:32
Air Blank	0.000	00:26:07

Preventative Maintenance
Perf time = Time Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J SANDALL 187 AGENCY LNC PD

DATE 9/2/16 TIME 0026

INTOXILYZER SERIAL # 30-001873 LOCATION MRIJ JRIL

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE J Sandall 187

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
09/02/2016

00:26:14

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP, TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:27:04
Diagnostic Test	Pass	00:27:39
Air Blank	0.000	00:28:06
0.100 Cal Check	0.097	00:28:27
Air Blank	0.000	00:28:56
Subject Test	0.000	00:29:32
Air Blank	ABT*	00:29:40
Air Blank	IPS**	00:29:47

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80 001873
Core Version: 8105.48
09/02/2016

00:39:31

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP, TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:40:28
Diagnostic Test	Pass	00:41:03
Air Blank	0.000	00:41:30
0.100 Cal Check	0.097	00:41:51
Air Blank	0.000	00:42:21
Subject Test	0.188*	00:45:28
Air Blank	0.000	00:46:01
Air Blank	0.000	00:46:28
Subject Test	XXX**	00:46:54
Air Blank	RFI***	00:47:06

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. S. Small 187 AGENCY LNCAD
DATE 9/2/16 TIME 0048
INTOXILYZER SERIAL # 80-001873 LOCATION MARICOPA JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. S. Small 187

187

INTOXILYZER 8000
Location: LHC MAIN JATI
Serial Number: 80 001873
Firmware Version: 8106.48
09/02/2016

00:48:19

Standard Lot#: 22315100A1
Last Changed By: J. SONDRON #187

QAS: J. SONDRON #187
LAKE HAVASU CITY PD

Operator: J. SONDRON #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	00:48:48
0.100 mL Check	0.097	00:49:00
Air Blank	0.000	00:49:33

Preventative Maintenance
Performed: Timer Reset