



POLICE DEPARTMENT  
 LAKE HAVASU CITY  
 2360 McCULLOCH BOULEVARD NORTH  
 LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Louis H. Plunkett III is a Certified Intoxifyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedures are full, a true, and correct copy of the original.

 IF 253  
 \_\_\_\_\_  
 Signature

STATE OF ARIZONA        )  
   ) ss  
 COUNTY OF MOHAVE     )


Subscribed and sworn to before me this   4   day of   July  .

  2018   by   Ashley Rae Bonds  

  
 \_\_\_\_\_  
 NOTARY PUBLIC

  7/4/18    
 \_\_\_\_\_  
 DATE

Commission expires:   May, 29 2021  



Area Code 928

Administration .....	855-4884	Fax #'s		Emergency .....	9-1-1
Business Office .....	855-1171	Administration .....	680-5430	Non-Emergency Dispatch/ .....	855-4111
Investigations .....	855-5775	Business Office .....	680-5431	Havasu Silent Witness .....	854-TIPS
Patrol .....	855-0515	Investigations/Patrol .....	680-5432		

E-mail: police@lhcaz.gov

Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)  
ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS H. PLUNKETT JR AGENCY LAKE HAVASU CITY PD

DATE 06-30-18 TIME 0918

INTOXILYZER SERIAL # 80-001029 LOCATION ROTARY PARK

(X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

(X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(X) 4. Air blank completed.

(X) 5. Calibration check completed. Test results 0.099 AC.

(X) 6. Air blank completed.

(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

Preventative Maintenance  
Performed -- Timer Reset

Exhibit G-4. Standard Quality Assurance Procedure Intoxilyzer, Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES

INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOUIS H. PLUNKETT III AGENCY WAYNE HANASHI CITY PD

DATE 06-30-18 TIME 0932

INTOXILYZER SERIAL # 80-001029 LOCATION ROTARY PARK

(X) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

(X) 1. Clock time check.

(X) 2. Date check.

OPERATIONAL TESTS

(X) 1. Alcohol-free subject test result 0.000 AC.

(X) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed

(X) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.

(X) 4. Standard Calibration Check standard 0.100 AC. Result 0.100 AC.

Instrument is operating properly and accurately. YES X NO

COMMENTS

SIGNATURE [Signature] 253

DPS Form Exh G-4 (Rev 05-01)

Historical Note

New Exhibit G-4 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).



Exhibit G-2. Standard Calibration Check Procedure, Intoxilyzer Model 8000

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)  
ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS W. PLUNKETT III AGENCY LAKE HAVASU CITY PD

TIME 0937

INTOXILYZER SERIAL # 80-001029 LOCATION ROTARY PARK

(X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C

(X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(X) 4. Air blank completed.

(X) 5. Calibration check completed. Test results 0. 098 AC.

(X) 6. Air blank completed.

(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  253

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

