

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDELL 187 AGENCY LNC PD

DATE 11/26/17 TIME 0121

INTOXILYZER SERIAL # 80-001873 LOCATION MARIN JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.037 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondell 187

INTOXILYZER 8000

Location: LHC MAIN JATT.

Serial Number: 80-001873

Core Version: 8105.48

01/26/2017

01:21:21

Standard Lot#: 22315100A1

Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187

LAKE HAVASU CITY PD

Operator: J. SONDROL #187

LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	01:21:52
0.100 Cal Check	0.097	01:22:12
Air Blank	0.000	01:22:42

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J Somdell 187 AGENCY LHC PD

DATE 1/26/17 TIME 0128

INTOXILYZER SERIAL # 80-001873 LOCATION MAIN JAIL

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE J Somdell 187

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
01/26/2017

01:22:59

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP,TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:23:45
Diagnostic Test	Pass	01:24:20
Air Blank	0.000	01:24:47
0.100 Cal Check	0.097	01:25:08
Air Blank	0.000	01:25:37
Subject Test	0.000	01:25:54
Air Blank	ABT*	01:26:00
Air Blank	IPS**	01:26:08

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

9157

INTOXILYZER 8000
Location: LHC MAIN JAIL
Serial Number: 80-001873
Core Version: 8105.48
01/26/2017

01:26:34

Standard Lot#: 22315100A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Subject: SQAP, TEST
DOB: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	01:27:19
Diagnostic Test	Pass	01:27:54
Air Blank	0.000	01:28:21
0.100 Cal Check	0.097	01:28:42
Air Blank	0.000	01:29:11
Subject Test	0.116*	01:32:19
Air Blank	0.000	01:32:51
Air Blank	0.000	01:33:18
Subject Test	XXX**	01:33:54
Air Blank	RFI***	01:34:03

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDRA '18 AGENCY LNC PD
DATE 11/24/17 TIME 0135
INTOXILYZER SERIAL # 82-001873 LOCATION MOBILE JAIL

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .102 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.096 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000

Location: LHC MAIN JAIL

Serial Number: 80-001873

Core Version: 8105.48

01/26/2017

01:35:10

Standard Lot#: 22315100A1

Last Changed By: J. SONDRON #187

QAS: J. SONDRON #187

LAKE HAVASU CITY PD

Operator: J. SONDRON #187

LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	01:35:39
0.100 Cal Check	0.096	01:35:59
Air Blank	0.000	01:36:29

Preventative Maintenance
Performed -- Timer Reset