



POLICE DEPARTMENT

2360 McCULLOCH BOULEVARD NORTH
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Josh Sondrol is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Josh Sondrol
Signature

STATE OF ARIZONA)
) ss
COUNTY OF MOHAVE)

Subscribed and sworn to before me this 27th day of June.

2012 by Joshua Sondrol

Nileena Detha Rebecca Hinkle
NOTARY PUBLIC

6/27/12
DATE

Commission expires:

Area Code 928

Administration 865 4884
Records Office 865 4174
Investigations 865 5774
Patrol 865 0515

Fax #s
Administration 680 5430
Business Office 680 5431
Investigations/Patrol 680 5432

Emergency 911
Non-Emergency Dispatch 865 4111
Human Talent/Witness 865 4100

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LNC PD
DATE 6/27/17 TIME 0909
INTOXILYZER SERIAL # 80-003355 LOCATION CHANNEL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

187

SIGNATURE J Sondrol 187

INTOXILYZER 8000
Instrument Initialization
08:39:37 06/27/2017

INTOXILYZER 8000
Location: CHANNEL BOOKING
Serial Number: 80-003355
Core Version: 8105.48
06/27/2017 09:09:41

Standard Lot#: 2231510CA1
Last Changed By: J. SONDROL #187

CAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/210L	Time
Air Blank	0.000	09:10:13
0.100 Cal Check	0.100	09:10:34
Air Blank	0.000	09:11:13

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J Sondrol 187 AGENCY LHC PD
DATE 6/27/17 TIME 0900
INTOXILYZER SERIAL # 80-003355 LOCATION CHANNEL

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

INTOXILYZER 8000
Location: CHANNEL BOOKING
Serial Number: 80-103355
Core Version: 8105.48
06/27/2017 09:12:55

Standard Lot#: 2231510091
Last Changed By: J. SONDRCL #187

QAS: J. SONDRCL #187
LAKE HAVASU CITY PD

Operator: J. SONDRCL #187
LAKE HAVASU CITY PD

Subject: SGAP, TEST
DC3: 01/01/1950
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

SIGNATURE J Sondrol 187

DPS Form Exh G-4 (Rev 05-01)

INSTRUMENT ID:	06/27/2017	
LOCATION:	CHANNEL BOOKING	
SERIAL NUMBER:	80-103355	
CORE VERSION:	8105.48	
STANDARD LOT#:	2231510091	
LAST CHANGED BY:	J. SONDRCL #187	
QAS:	J. SONDRCL #187	
OPERATOR:	J. SONDRCL #187	
LCFPD		
SUBJECT:	SGAP, TEST	
DOB:	01/01/1950	
SEX:	M Weight: 123	
15 Minute Deprivation Period?	Yes	
Test	g/210L	Time
Air Blank	0.00	09:21:04
Diagnostic Test	Pass	09:21:42
Air Blank	0.00	09:22:39
0.100 Cal Check	0.100	09:22:30
Air Blank	0.00	09:22:59
Subject Test:	0.162*	09:26:07
Air Blank	0.00	09:26:40
Air Blank	0.00	09:27:08
Subject Test	0.00	09:27:37
Air Blank	0.00	09:27:45

*Deficient Sample
**Invalid Sample
Begin 15k deprivation period
***RF: Detect

Test	g/210L	Time
Air Blank	0.00	09:13:40
Diagnostic Test	Pass	09:14:16
Air Blank	0.00	09:14:45
0.100 Cal Check	0.099	09:15:06
Air Blank	0.100	09:15:35
Subject Test	0.001	09:15:52
Air Blank	0.00	09:15:58
Air Blank	0.00	09:16:36

*Sequence Aborted
**Inappropriate Sample

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDROL 187 AGENCY LINC PIS
DATE 6/27/17 TIME 0928
INTOXILYZER SERIAL # 80-003355 LOCATION CHANNEL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.096 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondrol 187

DPS Form Exh G-2 (Rev 05-01)

INTOXILYZER 8000
Location: CHANNEL BOOKING
Serial Number: 80-003355
Core Version: 6105.48
06/27/2017

J.S.

09:28:11

Standard Lot#: 22315109A1
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187
LAKE HAVASU CITY PD

Operator: J. SONDROL #187
LAKE HAVASU CITY PD

Test	g/21.0L	Time
Air Blank	0.000	09:28:40
0.100 Cal Check	1.096	09:29:01
Air Blank	0.000	09:29:31

Preventative Maintenance
Performed -- Timer Reset