

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SONDRAL 187 AGENCY LHC PD  
DATE 8/22/17 TIME 2315  
INTOXILYZER SERIAL # 80-003356 LOCATION MAIN JAIL

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sondral 187

DPS Form Exh G-2 (Rev 05-01)

COMPLETED 2 DAYS LATE DUE TO VACATION

187

INTOXILYZER 8000  
Location: MAIN JAIL  
Serial Number: 80-003356  
Core Version: 8105.48  
08/22/2017 23:15:57

-----  
Standard Lot#: 22315100A1  
Last Changed By: J. SONDROL #187

-----  
QAS: J. SONDROL #187  
LAKE HAVASU CITY PD

-----  
Operator: J. SONDROL #187  
LAKE HAVASU CITY PD

-----

Test	g/210L	Time
Air Blank	0.000	23:16:28
0.100 Cal Check	0.098	23:16:48
Air Blank	0.000	23:17:18

Preventative Maintenance  
Performed -- Timer Reset