



EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J SANDRICK 187 AGENCY LHC PD

DATE 6/23/17 TIME 0045

INTOXILYZER SERIAL # 80-003356 LOCATION MAN JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J Sandrick 187

187

INTOXILYZER 8000  
Location: MAIN JAIL  
Serial Number: 80-003356  
Core Version: 8105.48  
06/23/2017 00:45:19

-----  
Standard Lot#: 22315100A1  
Last Changed By: J. SONDROL #187  
-----

QAS: J. SONDROL #187  
LAKE HAVASU CITY PD  
-----

Operator: J. SONDROL #187  
LAKE HAVASU CITY PD  
-----

Test	g/210L	Time
Air Blank	0.000	00:45:58
0.100 Cal Check	0.098	00:46:19
Air Blank	0.000	00:46:49

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J Smidel 187 AGENCY LUCAS

DATE 6/23/17 TIME 0048

INTOXILYZER SERIAL # 80-003356 LOCATION MAIN TRAIL

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE J Smidel 187

187

INTOXILYZER 8000  
Location: MAIN JAIL  
Serial Number: 80-003356  
Core Version: 8105.48  
06/23/2017 00:48:36

Standard Lot#: 22315100A1  
Last Changed By: J. SONDROL #187

QAS: J. SONDROL #187  
LAKE HAVASU CITY PD

Operator: J. SONDROL #187  
LAKE HAVASU CITY PD

Subject: SQAP,TEST  
DOB: 01/01/1950  
Sex: M Weight: 123

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:49:22
Diagnostic Test	Pass	00:49:57
Air Blank	0.000	00:50:24
0.100 Cal Check	0.097	00:50:45
Air Blank	0.000	00:51:15
Subject Test	0.000	00:51:42
Air Blank	ABT*	00:51:56
Air Blank	IPS**	00:52:04

\*Sequence Aborted  
\*\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: MAIN JAIL  
Serial Number: 80-003356  
Core Version: 8105.48  
06/23/2017

00:52:59

---

Standard Lot#: 22315100A1  
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187  
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187  
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST  
DOB: 01/01/1950  
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	00:53:43
Diagnostic Test	Pass	00:54:18
Air Blank	0.000	00:54:46
0.100 Cal Check	0.097	00:55:07
Air Blank	0.000	00:55:37
Subject Test	0.303*	00:58:44
Air Blank	0.000	00:59:17
Air Blank	0.000	00:59:45
Subject Test	0.025	01:00:25
Air Blank	RFI**	01:00:41
Air Blank	0.000	01:01:10

\*Deficient Sample

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: MAIN JAIL  
Serial Number: 80-003356  
Core Version: 8105.48  
06/23/2017 01:01:50

---

Standard Lot#: 22315100A1  
Last Changed By: J. SONDROL #187

---

QAS: J. SONDROL #187  
LAKE HAVASU CITY PD

---

Operator: J. SONDROL #187  
LAKE HAVASU CITY PD

---

Subject: SQAP,TEST  
DOB: 01/01/1950  
Sex: M Weight: 123

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	01:02:36
Diagnostic Test	Pass	01:03:11
Air Blank	0.000	01:03:39
0.100 Cal Check	0.097	01:03:59
Air Blank	0.000	01:04:29
Subject Test	XXX*	01:04:55
Air Blank	0.000	01:05:27

\*Invalid Sample  
Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. SONDRELL 187 AGENCY LNA PD  
DATE 6/23/17 TIME 0106  
INTOXILYZER SERIAL # 80-003356 LOCATION MARICOPA JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(✓) 4. Air blank completed.  
(✓) 5. Calibration check completed. Test results 0.097 AC.  
(✓) 6. Air blank completed.  
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]



INTOXILYZER 8000  
Location: MAIN JAIL  
Serial Number: 80-003356  
Core Version: 8105.48  
06/23/2017

01:06:21

-----  
Standard Lot#: 22315100A1  
Last Changed By: J. SONDROL #187  
-----

QAS: J. SONDROL #187  
LAKE HAVASU CITY PD  
-----

Operator: J. SONDROL #187  
LAKE HAVASU CITY PD  
-----

Test	g/210L	Time
Air Blank	0.000	01:06:50
0.100 Cal Check	0.097	01:07:11
Air Blank	0.000	01:07:40

Preventative Maintenance  
Performed -- Timer Reset