



POLICE DEPARTMENT  
**LAKE HAVASU CITY**  
2360 McCULLOCH BOULEVARD NORTH  
LAKE HAVASU CITY, ARIZONA 86403-5947

To Whom It May Concern:

I hereby attest that Louis H. Plunkett III is a Certified Intoxilyzer Quality Assurance Specialist for the Lake Havasu City Police Department.

I certify that the attached documentation of Standard Quality Assurance Procedures and/or Standard Calibration Check Procedure are full, true and correct copy of the original.

Signature

STATE OF ARIZONA        )  
  ) ss  
COUNTY OF MOHAVE    )

Subscribed and sworn to before me this 25 day of September 2017


\_\_\_\_\_ by Ashley Rae Bonds

NOTARY PUBLIC

9/25/17

DATE

Commission expires: May, 29 2021



Area Code 928

Administration ..... 855-4884  
Business Office ..... 855-1171  
Investigations ..... 855-5775  
Patrol ..... 855-0515

Fax #'s  
Administration ..... 680-5430  
Business Office ..... 680-5431  
Investigations/Patrol ..... 680-5432

Emergency ..... 9-1-1  
Non-Emergency Dispatch/ ..... 855-4111  
Havasu Silent Witness ..... 854-TIPS

E-mail: [police@lhcaz.gov](mailto:police@lhcaz.gov)

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS H. BLUNKETT III 263 AGENCY LHCPO

DATE 09/26/17 TIME 0706

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPO JAIL

(X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

(X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(X) 4. Air blank completed.

(X) 5. Calibration check completed. Test results 0. .100 AC.

(X) 6. Air blank completed.

(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 263



EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOUIS H. BLUNKET JR 253 AGENCY LHCPD

DATE 09/25/17 TIME 0713

INTOXILYZER SERIAL # 80-003356 LOCATION LHCPD JAIL

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

---

---

---

---

SIGNATURE  253

INTOXILYZER 8000  
Location: MAIN JAIL  
Serial Number: 80-003356  
Core Version: 8105.48  
09/25/2017

07:13:46

---

Standard Lot#: 13917100A1  
Last Changed By: J. SONDROL #187

---

QAS: L. PLUNKETT III #253  
LAKE HAVASU CITY PD

---

Operator: L. PLUNKETT III #253  
LAKE HAVASU CITY PD

---

Subject: SQAP, TEST  
DOB: 11/11/2001  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	07:15:01
Diagnostic Test	Pass	07:15:36
Air Blank	0.000	07:16:04
0.100 Cal Check	0.100	07:16:24
Air Blank	0.000	07:16:54
Subject Test	0.000	07:17:12
Air Blank	ABT*	07:17:18
Air Blank	IPS**	07:17:26

\*Sequence Aborted

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: MAIN JAIL  
Serial Number: 80-003356  
Core Version: 8105.48  
09/25/2017

07:18:18

---

Standard Lot#: 13917100A1  
Last Changed By: J. SONDROL #187

---

QAS: L. PLUNKETT III #253  
LAKE HAVASU CITY PD

---

Operator: L. PLUNKETT III #253  
LAKE HAVASU CITY PD

---

Subject: SQAP, TEST  
DOB: 11/11/2001  
Sex: M Weight: 111

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	07:19:09
Diagnostic Test	Pass	07:19:44
Air Blank	0.000	07:20:11
0.100 Cal Check	0.101	07:20:32
Air Blank	0.000	07:21:02
Subject Test	0.042*	07:24:09
Air Blank	0.000	07:24:41
Air Blank	0.000	07:25:08
Subject Test	XXX**	07:25:34
Air Blank	RFI***	07:25:44

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

\*\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOUIS W. PUNNETT III 253 AGENCY MACPD

DATE 09/26/17 TIME 0730

INTOXILYZER SERIAL # 80-003366 LOCATION MACPD JAIL

(X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$

(X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(X) 4. Air blank completed.

(X) 5. Calibration check completed. Test results 0.000 AC.

(X) 6. Air blank completed.

(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  253

INTOXILYZER 8000  
Location: MAIN JAIL  
Serial Number: 80-003356  
Core Version: 610s.46  
09/25/2007 07:30:10

-----  
Standard Lot#: 13917100A1  
Last Changed By: J. SONDROT #187

-----  
CAS: L. PLUNKETT III #253  
LAKE HAVASU CITY PD

-----  
Operator: L. PLUNKETT III #253  
LAKE HAVASU CITY PD  
-----

Test	g/2100	Time
Air Blank	0.000	07:30:47
0.100 Cal Check	0.099	07:31:02
Air Blank	0.000	07:31:31

Preventative Maintenance  
Performed: Timer Reset